

No. 29/MISC/3/2018-DC (165)
Government of India
Directorate General of Health Services
Ministry of Health & Family Welfare
Central Drugs Standard Control Organisation
(Medical Devices Division)


Date: 08.11.2019

Notice

Various stakeholders have sought a clarification regarding the environmental conditions for the equipments which are to be regulated from 01.01.2020 and 01.04.2020.

In this regard, it is to clarify that the environmental conditions referred in fifth schedule, annexure 'A' of Medical Devices Rules, 2017 such as well ventilated area with neat and clean environment, free from dust and other particulate matter, air conditioning will be applicable for such equipments.

A proposal will be moved to DTAB, stakeholders are requested to send their comments on ddcimnd-cdsco@nic.in


(Dr. V. G. Somani)
Drugs Controller General (I)