

भारत सरकार
स्वास्थ्य एवं परिवार कल्याण मंत्रालय
(स्वास्थ्य सेवा महानिदेशालय)
केन्द्रीय औषध मानक नियंत्रण संगठन
केन्द्रीय औषधि परीक्षण प्रयोगशाला
क्षेत्रीय अन्न व औषध प्रशासन भवन, जी.एम.एस.डी कंपाउंड,
बेलासीस रोड, मुंबई सेंट्रल, मुंबई - 400008.
दूरभाष : 022-23002309 /022-23002138.
फैक्स : 022-23099240.

ईमेल- : cdtlmumbai@cdsco.nic.in



CDSCO



Government of India
Ministry of Health and Family Welfare
(Directorate General of Health Services)
Central Drugs Standard Control Organisation
Central Drugs Testing Laboratory
Zonal FDA Bhawan, GMSD Compound,
Bellasis Road, Mumbai Central, Mumbai - 400008.
Tel : 022-23002309 /022-23002138.
Fax : 022-23099240.
Email : cdtlmumbai@cdsco.nic.in

Accredited By: NABL (ISO/IEC-17025:2017 only for scope under Chemical & Biological Testing)
Certified For: IMS (ISO - 9001:2015; ISO - 14001:2015; ISO - 45001:2018)

No. 1(31)/CDTL-MUM/24-25/ 511

Date: 14/05/24


CORRIGENDUM

Subject: Notice inviting tender (NIQ) No. 1(31)/CDTL-MUM/24-25/409, dt. 7/5/24 for Supply of GC Columns at Central Drugs Testing Laboratory, Mumbai.

The bid for the GC Columns having file number as mentioned above has been revised with respect to the particle size given below:

Sr. No.	Existing Clause	Revised Clause
2	a fused silica capillary column 30 m x 0.32mm x 0.3µm, coated with Polyethyelene Glycol (av Mol wt ab 15000)	a fused silica capillary column 30 m x 0.32mm x 2µm, coated with Polyethyelene Glycol (av Mol wt ab 15000)
3	a fused silica capillary column 30 m x 0.53mm x 0.3µm, coated with6% cyanopropyl/phenyl, 94% polydimethylsiloxane	a fused silica capillary column 30 m x 0.53mm x 3µm, coated with6% cyanopropyl/phenyl, 94% polydimethylsiloxane
4	a fused silica capillary column 30 m x 0.53mm x 0.3µm, coated with6% cyanopropyl/phenyl, 94% polydimethylsiloxane	a fused silica capillary column 30 m x 0.53mm x 3µm, coated with6% cyanopropyl/phenyl, 94% polydimethylsiloxane
6	a fused silica capillary column 30 m x 0.53mm x 0.3µm, coated with(50%-Phenyl)- methylpolysiloxane	a fused silica capillary column 30 m x 0.53mm x 0.5µm, coated with(50%-Phenyl)- methylpolysiloxane

Sealed quotations may be submitted by 10/06/2024 upto 3:00 PM and will be opened on 10/06/2024 at 3:30 PM. Other terms and conditions of Notice Inviting Quotation (NIQ) remain unchanged.


14.5.24
Sr. Scientific Officer,
For Director, Central Drugs Testing Laboratory,
Mumbai.