

**GOVERNMENT OF INDIA
MINISTRY OF HEALTH AND FAMILY WELFARE
DIRECTORATE GENERAL OF HEALTH SERVICES
NATIONAL TOBACCO TESTING LABORATORY
OFFICE OF THE REGIONAL DRUGS TESTING LABORATORY
SIX MILE, GUWAHATI-781022
Ph. 0361-2338555, FAX: 0361-2330555
Email: dirnttl.gwh-mohfw@gov.in**

No. NTTL/Ghy/Equipt/21/2023-24/ 539

Date: 04.01.2024


Notice Inviting Quotation

The Director-in-charge, National Tobacco Testing Laboratory, Guwahati invites quotation from the interested and bonafide Manufacturer, Dealer, Distributer or retailer providing and supplying of Capillary GC Column at National Tobacco Testing Laboratory, O/o Regional Drugs Testing Laboratory, Guwahati.

For detailed notice, please visit <https://www.cdsc.gov.in>

The detailed terms & condition of quotation for providing and supplying of Capillary GC Column is available on the above website. Sealed quotation, accepting the stipulated terms & condition should be sent by post so as to reach the National Tobacco Testing Laboratory, O/o Regional Drugs Testing Laboratory, Guwahati, Ministry of Health & Family Welfare, DGHS, Govt. of India, Six Mile, Guwahati-781022 by 23.01.2024 up to 3:00PM or dropped in the tender box on any working day between 11:00 AM to 3:00PM up to 23.01.2024 and Tender/Quotation will be opened on the same day at 3:30pm 23.01.2024 in presence of interested Quotioners.

Yours faithfully,



For Director in-charge,
National Tobacco Testing Laboratory,
Guwahati

Annexure-A

Sl. No.	Name of Items	Quantity
1	Capillary GC Column Stationary Phase: 6% Cyanopropylphenyl-94 30m x 0.53 μ m, Film – 3.0 μ m USP phase – G43	1 No.
2	Capillary GC Column Stationary Phase: CP Wax 51 or equivalent (base deactivated PEG) 25 - 30m x 0.25mm Film – 0.2 μ m	1 No.

PSha
04/01/2024

**GOVERNMENT OF INDIA
MINISTRY OF HEALTH AND FAMILY WELFARE
DIRECTORATE GENERAL OF HEALTH SERVICES
NATIONAL TOBACCO TESTING LABORATORY
OFFICE OF THE REGIONAL DRUGS TESTING LABORATORY
SIX MILE, GUWAHATI-781022
Ph. 0361-2338555, FAX: 0361-2330555
Email: dirnttl.gwh-mohfw@gov.in**

No. NTTL/Ghy/Equipt/21/2023-24/ 539

Date: 04.01.2024


NOTICE INVITING QUOTATION

Quotation is hereby invited from the interested and bonafide Manufacturer, Dealer, Distributer or retailer providing and supplying of Capillary GC Column at National Tobacco Testing Laboratory, O/o Regional Drugs Testing Laboratory, Guwahati under the following specification, terms and conditions:

Sl. No.	Name of Product Description	Quantity
1	Capillary GC Column Stationary Phase: 6% Cyanopropylphenyl-94 30m x 0.53µm, Film – 3.0 µm USP phase – G43	1 No.
2	Capillary GC Column Stationary Phase: CP Wax 51 or equivalent (base deactivated PEG) 25 - 30m x 0.25mm Film – 0.2 µm	1 No.

N.B: Applicable Taxes and duties if any, which may be levied by the Government, should be shown separately.

Yours faithfully,



For Director in-charge,
National Tobacco Testing Laboratory,
Guwahati


Terms & Conditions:

1. The quotation/tender should be submitted in sealed covers. Open quotation/tender will be summarily rejected.
2. The quotation should be valid for a period of 90 (ninety) days from the date of opening of the tender / quotation.
3. The Director-in-Charge, National Tobacco Testing Laboratory, Guwahati reserves the right to reject the quotation or supply order at any stage without assigning any reasons thereof.
4. The quoted rate should be all inclusive and nothing will be paid extra for transportation, loading, un-loading at National Tobacco Testing Laboratory, Guwahati.
5. The Quotationer / supplier should supply of Capillary GC Column in good condition, recently manufactured & strictly as per specification within thirty (30) days from the date of issue of supply order.
6. The person duly authorized by the quotationer will be allowed to deliver Capillary GC Column at National Tobacco Testing Laboratory, Guwahati.
7. No "Way Bill" in connection with supply of Capillary GC Column will be issued from National Tobacco Testing Laboratory, Guwahati.
8. The quotationer should clearly mention the name of manufacturer, Catalogue No./Product No. and year of manufacturing.
9. That on successful completion of supply of Capillary GC Column per specification, make and Catalogue Number to Store Section, National Tobacco Testing Laboratory, Guwahati, the quotationer / supplier should submit bill in triplicate along with mandate form to release e-payment through PFMS.
10. Only manufacturing firms / authorized distributors / retailer are eligible to participate.
11. The quotationer should clearly quote their rates for both in figures and in words in net price (including any other taxes/discounts etc.) failing which your quotation will not be entertained. Any overwriting in rates or any change or alteration without attestation is liable to eliminate in the outright rejection of the quotation. In case there is any error between the amount specified in figures and words, the lower of the two will be taken for consideration.
12. The quotation duly completed as per the instruction given herein and accepting the same may be sent by speed post or dropped in the "Tender Box" on any working days between 11:00AM to 3:00PM up to 23.01.2024 at National Tobacco Testing Laboratory, Office of Regional Drugs Testing Laboratory, Guwahati, Ministry of Health & Family Welfare, DGHS, Govt. of India, Six Mile, Guwahati-781022.
"Quotation for supply of Capillary GC Column" and Quotation "No.

NTTL/Ghy/Equipt/21/2023-24/539 dt. 04.01.2024" should be written on the top of the envelop.

13. Late quotation i.e. quotation received after the specified date and time of receipt will not be considered in any circumstances.
14. The quotation so received will be opened on the same day i.e. on 23.01.2024 at 3:30PM in presence of interested Quotationer (s) or their authorized representative.

Yours faithfully,



For Director in-charge,
National Tobacco Testing Laboratory,
Guwahati