

No . 6-3(Store)/2020/ 2588
OFFICE OF THE DIRECTOR
CENTRAL DRUGS LABORATORY,
GOVERNMENT OF INDIA
MIN. OF HEALTH & FAMILY WELFARE
3, KYD STREET, KOLKATA-16

Kolkata Dated the 12/05/2020


NOTICE INVITING QUOTATION

The Director-in Charge, Central Drugs Laboratory, Kolkata invites quotation from the interested and bonafide Manufacturer, Dealer, Distributer or retailer providing and supplying of **Monochromator Based UV-Visible Microplate Reader** : at Central Drugs Laboratory, Kolkata.

For detailed Notice. please visit <http://www.cdsc.gov.in>

The detailed terms & condition of quotation for providing and supplying of **Monochromator Based UV-Visible Microplate Reader** is available on the above website. Filled in quotation, accepting the stipulated terms & condition may be dropped in tender box available at the wing of Jr. Administrative Officer, Central Drugs Laboratory, Kolkata on any working day upto **3:00pm on 16.06.2020** and Tender / Quotation will be opened on the same day at **3:30pm on 16.06.2020** in presence of interested Quotationers.

Yours faithfully


(Dr. Anjan Pal) 11/05/2020
SSO (Microbiology)
for Director
Central Drugs Laboratory, Kolkata.

No. 6-3/2020-S/2528
OFFICE OF THE DIRECTOR
CENTRAL DRUGS LABORATORY,
GOVERNMENT OF INDIA
MIN. OF HEALTH & FAMILY WELFARE
3, KYD STREET, KOLKATA-16

Kolkata Dated the 12/05/2020

NOTICE INVITING QUOTATION

Quotation is hereby invited from the interested and bona fide Dealer, Distributers, Retailers for Providing and supplying of Monochromator Based UV-Visible Microplate Reader under the following specification, terms and conditions:

Make : Thermo Fisher / Berthold / Biorad

Monochromator Based UV-Visible Microplate Reader :

1 No.

Technical Specification :

1. Instrument should have scanning unit for UV-Visible wavelength range, with dedicated cuvette port and should be able to read 96 and 384 microwell plate format.
2. Instrument is able to provide the Wavelength range is about from 200nm to 950nm with 1nm steps.
3. Wavelength Selection Monochromator & Filter
4. Precision : Better 2% or lesser over entire range
5. Accuracy : Better 2% or lesser over entire range
6. Should be able to read end point, kinetics, spectral scanning and also kinetic spectral scan
7. Performance Specifications : Bandwidth : <2.5nm or better and Xenon flash lamp
8. Instrument should have inbuilt incubation and linear shaking options for longer kinetics assays.
9. Incubation Temperature : 5°C to 45°C ±3°C
10. The instrument should have a memory of not less than 50 inbuilt protocols in stand-alone mode
11. Software should be CFR21 Part 11 regulation compliance
12. Instrument should have suitable Data Analysis Software
13. Preferably Instrument should be operated through intuitive touch screen and with external PC.
14. Warranty preferably 2 years
15. Onsite validation IQ/OQ/PQ
16. Vendor should ensure at least 7 years service backup along with parts availability of the quoted model
17. External PC along with suitable printer should be provided with the instrument (optional)

N.B: Applicable Taxes and duties if any, which may be levied by the Government, should be shown separately.

Yours faithfully



Dr. Anjan Pal, 11/05/2020
SSO (Microbiology)

for Director


Central Drugs Laboratory, Kolkata.

Page 3 of 4

Terms & Conditions. :

1. The quotation/tender should be submitted in sealed covers. Open quotation/tender will be summarily rejected.
2. The quotation should be valid for a period of 90 (ninety) days from the date of opening of the tender / quotation.
3. The Director-in-Charge, CDL, Kolkata reserves the right to reject the quotation and/or supply order at any stage without assigning any reasons thereof.
4. The quoted rate should be all inclusive and nothing will be paid extra for transportation, loading, unloading at Central Drugs Laboratory, Kolkata.
5. The Quotationer/supplier should supply of **Monochromator Based UV-Visible Microplate Reader** in good condition & strictly as per specification within thirty (30) days from the date of issue of supply order.
6. The person duly authorized by the quotationer will be allowed to deliver **Monochromator Based UV-Visible Microplate Reader** at Central Drugs Laboratory, Kolkata on any working day within the time schedule.
7. No "Way Bill" in connection with supply of **Monochromator Based UV-Visible Microplate Reader** will be issued from Central Drugs Laboratory, Kolkata.
8. **The quotationer should clearly mention the name of manufacturer / Catalogue No. / Product No. and year of manufacturing.**
9. That on successful completion of supply of **Monochromator Based UV-Visible Microplate Reader** as per specification, make and Catalogue Number to Store Section, CDL, Kolkata, the quotationer/supplier should submit bill in triplicate along with mandate form to release e-payment through Pay & Accounts Office, Min. of Health & F.W. Kolkata.
10. **Only manufacturing firms / authorized distributors are eligible to participate.**
11. The Quotationer should clearly quote their rates for both in figures and in words **in net price (including any other taxes/discounts etc.) failing which your quotation will not be entertained.** Any overwriting in rates or any change and/or alteration without attestation is liable to culminate in the outright rejection of the quotation. In case there is any error between the amount specified in figures and words, the lower of the two will be taken for consideration.
12. The Quotation duly completed as per the instruction given herein and accepting the same may be dropped in the "Tender Box" on any working days between **11.00 AM to 3.00 P.M. up to** at the wing of Jr. Administrative Officer, Central Drugs Laboratory, Government of India, Ministry of Health & Family Welfare, 3, Kyd Street, Kolkata-700016.
13. Late quotation i.e. quotation received after the specified date and time of receipt will not be considered in any circumstances.
14. The Quotation so received will be opened on the same day i.e. on **16.06.2020 at 3.30 P.M.** in presence of interested Quotationer(s) or their authorized representative.

Yours faithfully


Dr. Anjan Pal,
SSO (Microbiology)
for Director

Central Drugs Laboratory, Kolkata.