

No . 6-3/2019-S/ 2368
OFFICE OF THE DIRECTOR
CENTRAL DRUGS LABORATORY,
GOVERNMENT OF INDIA
MIN. OF HEALTH & FAMILY WELFARE
3, KYD STREET, KOLKATA-16

Kolkata Dated the 02/03/20.


NOTICE INVITING QUOTATION

The Director-in Charge, Central Drugs Laboratory, Kolkata invites quotation from the interested and bonafide Manufacturer, Dealer, Distributer or retailer providing and supplying of **HPLC Column** at Central Drugs Laboratory, Kolkata.

For detailed Notice. please visit <http://www.cdsc.gov.in>

The detailed terms & condition of quotation for providing and supplying of **HPLC Column** is available on the above website. Filled in quotation, accepting the stipulated terms & condition may be dropped in tender box available at the wing of Jr. Administrative Officer, Central Drugs Laboratory, Kolkata on any working day upto **3:00pm on 17.03.2020** and Tender / Quotation will be opened on the same day at **3:30pm on 17.03.2020** in presence of interested Quotationers.

Yours faithfully


02/03/2020

(Dr. Anjan Pal)
SSO (Microbiology)
for Director
Central Drugs Laboratory, Kolkata.

Kolkata Dated the 02/03/2020


Quotation is hereby invited from the interested and bona fide Dealer, Distributors, Retailers for Providing and supplying of HPLC Column under the following specification, terms and conditions:

Make : Waters / Agilent / Thermo Fisher / Phenomenex / YMC / ACE

1. **HPLC L17 Column :**
 Stainless Steel Column 30cm x 7.8mm, 7µm, packed with strong cation-exchange resin consisting of sulphonated cross-linked styrene-divinylbenzene copolymer in the hydrogen form and a guard column 4cmx7.8mm packed with the same column material : 1 No.
2. **HPLC L26 Column :**
 Stainless Steel Column 25cm x 4.6mm, 5µm, packed with Butylsilyl Silicagel : 1 No.
3. **HPLC L59 Column :**
 Stainless Steel Column 30cm x 7.8mm, 5µm, packed with hydrophilic silica gel of a grade suitable for fractionation of globular proteins in the relative molecular mass range of 10000 to 50000 : 1 No.
4. **HPLC C18 Column :**
 - i) Size : 4.6mm x 250mm, 5µm : 6 Nos.
 - ii) Carbon Loading : Not less than 10%
 - iii) pH Range : 2.0 to 9.0 or better
 - iv) Temperature Limit : Not less than 70°C
 - v) Pore Size : ≤100 Å
 - vi) Surface Area : Not less than 200m²/g
5.
 - i) Size : 4.6mm x 150mm, 5µm : 4 Nos.
 - ii) Carbon Loading : Not less than 10%
 - iii) pH Range : 2.0 to 9.0 or better
 - iv) Temperature Limit : Not less than 70°C or above
 - v) Pore Size : ≤100 Å
 - vi) Surface Area : Not less than 200m²/g
6.
 - i) Size : 4.6mm x 100mm, 5µm - 3 Nos.
 - ii) Carbon Loading : Not less than 10%
 - iii) pH Range : 2.0 to 9.0 or better
 - iv) Temperature Limit : Not less than 70°C or above
 - v) Pore Size : ≤100 Å
 - vi) Surface Area : Not less than 200m²/g
7.
 - i) Size : 4.6mm x 100mm, 3µm - 1 No. (Nuclosil)
 - ii) Carbon Loading : Not less than 15%
 - iii) Should be endcapped
 - iv) pH Range : 2.0-8.0 or better
 - v) Temperature Limit : 70°C or above
 - vi) Pore Size : ≤100 Å
 - vii) Surface Area : Not less than 200m²/g
8.
 - i) Size : 3.9mm x 300mm, 5µm - 1 No.
 - ii) Carbon Loading : 18 to 25%
 - iii) pH Range : 1.0 to 12.0
 - iv) Pore Size : ≤100 Å
 - v) Surface Area : Not less than 200m²/g
9.
 - i) Size : 4.6mm x 150mm, 5µm : 3 Nos.
 - ii) Carbon Loading : 18 to 25%
 - iii) pH Range : 2.0 to 10.0
 - iv) Pore Size : ≤100 Å
 - v) Surface Area : Not less than 200m²/g
10.
 - i) Size : 4.6mm x 250mm, 5µm : 3 Nos.
 - ii) Carbon Loading : 11%
 - iii) pH Range : 1.0 to 12.0
 - iv) Temperature Limit : Not less than 70°C
 - v) Pore Size : ≤100 Å
 - vi) Surface Area : Not less than 200m²/g
11. **HPLC L51 Column :**
 - ii) Stainless Steel Column 25cm x 4.6mm packed with Amylose tris-3,5-dimethylphenylcarbamate-coated, porous, spherical, silica particles (10µm) available as Chiralpak AD from Chiral Technologies : 1 No.

Terms & Conditions. :

1. The quotation/tender should be submitted in sealed covers. Open quotation/tender will be summarily rejected.
2. The quotation should be valid for a period of 90 (ninety) days from the date of opening of the tender / quotation.
3. The Director-in-Charge, CDL, Kolkata reserves the right to reject the quotation and/or supply order at any stage without assigning any reasons thereof.
4. The quoted rate should be all inclusive and nothing will be paid extra for transportation, loading, unloading at Central Drugs Laboratory, Kolkata.
5. The Quotationer/supplier should supply of **HPLC Column** in good condition & strictly as per specification within thirty (30) days from the date of issue of supply order.
6. The person duly authorized by the quotationer will be allowed to deliver **HPLC Column** at Central Drugs Laboratory, Kolkata on any working day within the time schedule.
7. No "Way Bill" in connection with supply of **HPLC Column** will be issued from Central Drugs Laboratory, Kolkata.
8. **The quotationer should clearly mention the name of manufacturer and Product No.**
9. That on successful completion of supply of **HPLC Column** as per specification, make and Catalogue Number to Store Section, CDL, Kolkata, the quotationer/supplier should submit bill in triplicate along with mandate form to release e-payment through Pay & Accounts Office, Min. of Health & F.W. Kolkata.
10. **Only manufacturing firms / authorized distributors are eligible to participate.**
11. The Quotationer should clearly quote their rates for both in figures and in words **in net price (including any other taxes/discounts etc.) failing which your quotation will not be entertained.** Any overwriting in rates or any change and/or alteration without attestation is liable to culminate in the outright rejection of the quotation. In case there is any error between the amount specified in figures and words, the lower of the two will be taken for consideration.
12. The Quotation duly completed as per the instruction given herein and accepting the same may be dropped in the "Tender Box" on any working days between **11.00 AM to 3.00 P.M. up to 17.03.2020** at the wing of Jr. Administrative Officer, Central Drugs Laboratory, Government of India, Ministry of Health & Family Welfare, 3, Kyd Street, Kolkata-700016.
13. Late quotation i.e. quotation received after the specified date and time of receipt will not be considered in any circumstances.
14. The Quotation so received will be opened on the same day i.e. on **17.03.2020 at 3.30 P.M.** in presence of interested Quotationer(s) or their authorized representative.


Dr. Anjan Pal, 02/03/2020
SSO (Microbiology)
for Director
Central Drugs Laboratory, Kolkata.