

No . 6-4(Store)/2019/ 1315
OFFICE OF THE DIRECTOR
CENTRAL DRUGS LABORATORY,
GOVERNMENT OF INDIA
MIN. OF HEALTH & FAMILY WELFARE
3, KYD STREET, KOLKATA-16

Kolkata Dated the 10/10/19


NOTICE INVITING QUOTATION

The Director-in Charge, Central Drugs Laboratory, Kolkata invites quotation from the interested and bonafide Manufacturer, Dealer, Distributer or retailer providing and supplying of **Microprocessor Controlled Bench-top Refrigerated Centrifuge** at Central Drugs Laboratory, Kolkata.

For detailed Notice. please visit <http://www.cdscn.nic.in>

The detailed terms & condition of quotation for providing and supplying of **Microprocessor Controlled Bench-top Refrigerated Centrifuge** is available on the above website. Filled in quotation, accepting the stipulated terms & condition may be dropped in tender box available at the wing of Jr. Administrative Officer, Central Drugs Laboratory, Kolkata on any working day upto **3:00pm on 07.11.2019** and Tender / Quotation will be opened on the same day at **3:30pm on 07.11.2019** in presence of interested Quotationers.

Yours faithfully



(Dr. Anjan Pal)

SSO (Microbiology)

for Director

Central Drugs Laboratory, Kolkata.

No. 6-4/2019-S/ 1315
OFFICE OF THE DIRECTOR
CENTRAL DRUGS LABORATORY,
GOVERNMENT OF INDIA
MIN. OF HEALTH & FAMILY WELFARE
3, KYD STREET, KOLKATA-16

Kolkata Dated the 10/10/19

NOTICE INVITING QUOTATION

Quotation is hereby invited from the interested and bona fide Dealer, Distributors, Retailers for Providing and supplying of Bench Top Centrifuge under the following specification, terms and conditions:

Name of the Item	Quantity
<p>Microprocessor Controlled Bench-top Refrigerated Centrifuge :</p> <p><u>Make : Sigma / Thermo Scientific / Eppendorf</u></p> <p>i) System should be Microprocessor controlled</p> <p>ii) Maximum speed not less than 14000 RPM or 20000 x g RCF</p> <p>iii) Temperature Range - 10°C to 40°C (approx)</p> <p>iv) Runtime range 1 minute to 99 hours or better range with continuous operation mode</p> <p>v) System should have pulse / short run option</p> <p>vi) System should have digital display (LCD / TFT) for Speed, Time, RCF and Temperature</p> <p>vii) Should have Imbalance Detection System</p> <p>viii) System should be maintained with Direct and Brushless Induction Drive</p> <p>ix) The System Chamber should be made of Stainless Steel</p> <p>x) System should have CFC free Pre Cooling System</p> <p>xi) The system should be certified for quality and safety by International Certification Agency like CE marked and IVD Compliant</p> <p>xii) Rotor should have auto locking system and shall be installed and removed without requirement of any tools</p> <p>xiii) Power supply suitable for 230 V, AC Single Phase, 50 Hz operation</p> <p>xiv) Warranty : Minimum 1 year</p> <p>xv) Rotor :</p> <ul style="list-style-type: none">• Must have swinging out Rotor with capacity not less than 4x250±20 having maximum speed not less than 4500 RPM or 3200 x g RCF with adapter for 200ml±20ml tubes.• Fixed angle rotor with minimum 6x50 ml tubes having maximum speed not less than 10000 RPM or 14000 x g RCF with adapter for 15 ml tubes• Fixed angle rotor with minimum 24x1.5/2 ml tubes having maximum speed not less than 12000 RPM or 20000 x g RCF	1 No.

N.B: Applicable Taxes and duties if any, which may be levied by the Government, should be shown separately.

Yours faithfully



Dr. Anjan Pal,
SSO (Microbiology)
for Director

Central Drugs Laboratory, Kolkata.

Terms & Conditions. :

1. The quotation/tender should be submitted in sealed covers. Open quotation/tender will be summarily rejected.
2. The quotation should be valid for a period of 90 (ninety) days from the date of opening of the tender / quotation.
3. The Director-in-Charge, CDL, Kolkata reserves the right to reject the quotation and/or supply order at any stage without assigning any reasons thereof.
4. The quoted rate should be all inclusive and nothing will be paid extra for transportation, loading, unloading at Central Drugs Laboratory, Kolkata.
5. The Quotationer/supplier should supply of **Microprocessor Controlled Bench-top Refrigerated Centrifuge** in good condition & strictly as per specification within thirty (30) days from the date of issue of supply order.
6. The person duly authorized by the quotationer will be allowed to deliver **Microprocessor Controlled Bench-top Refrigerated Centrifuge** at Central Drugs Laboratory, Kolkata on any working day within the time schedule.
7. No "Way Bill" in connection with supply of **Microprocessor Controlled Bench-top Refrigerated Centrifuge** will be issued from Central Drugs Laboratory, Kolkata.
8. **The quotationer should clearly mention the name of manufacturer and Product No.**
9. That on successful completion of supply of **Microprocessor Controlled Bench-top Refrigerated Centrifuge** as per specification, make and Catalogue Number to Store Section, CDL, Kolkata, the quotationer/supplier should submit bill in triplicate along with mandate form to release e-payment through Pay & Accounts Office, Min. of Health & F.W. Kolkata.
10. **Only manufacturing firms / authorized distributors are eligible to participate.**
11. The Quotationer should clearly quote their rates for both in figures and in words **in net price (including any other taxes/discounts etc.) failing which your quotation will not be entertained.** Any overwriting in rates or any change and/or alteration without attestation is liable to culminate in the outright rejection of the quotation. In case there is any error between the amount specified in figures and words, the lower of the two will be taken for consideration.
12. The Quotation duly completed as per the instruction given herein and accepting the same may be dropped in the "Tender Box" on any working days between **11.00 AM to 3.00 P.M. up to 07.11.2019** at the wing of Jr. Administrative Officer, Central Drugs Laboratory, Government of India, Ministry of Health & Family Welfare, 3, Kyd Street, Kolkata-700016.
13. Late quotation i.e. quotation received after the specified date and time of receipt will not be considered in any circumstances
14. The Quotation so received will be opened on the same day i.e. on **07.11.2019 at 3.30 P.M.** in presence of interested Quotationer(s) or their authorized representative.

Yours faithfully



Dr. Anjan Pal,
SSO (Microbiology)
for Director

Central Drugs Laboratory, Kolkata.