

भारत सरकार

क्षेत्रीय औषध परीक्षण प्रयोगशाला

केन्द्रीय औषध मानक नियन्त्रण संगठन

(स्वास्थ्य सेवाओं का महानिदेशालय)

स्वास्थ्य एवं परिवार कल्याण मंत्रालय

सैक्टर-39 सी, चन्डीगढ़-160036

फोन नं: 0172-2688239, फैक्स नं: 0172-2636316



सत्यमेव जयते

NABL

ACCREDITED LABORATORY

GOVERNMENT OF INDIA

REGIONAL DRUGS TESTING LABORATORY

CENTRAL DRUGS STANDARD CONTROL ORGANISATION
(Directorate General of Health services)

Ministry of Health & Family Welfare

SECTOR-39 C, CHANDIGARH – 160036

Tel. No: -0172-2688239, Fax- No: 0172-2636316

E-mail id: rdtlchd@cdsco.nic.in

No. 3-8 /2018-19/St/Equipments/2430 /18

Dated: 17.08.2018

Tender notice

Subject:-Tender Rate Enquiry for supply of Vacuum Manifold with flask(6 Places each) in RDTL, Chandigarh.

Sealed Tender Rate Enquiries are invited from reputed interested and manufacturing firm(s)/distributors of Vacuum Manifold with flask (6 Places each) in RDTL, who have been rendering/undertaking the specified work in the field of Scientific test equipment as per the specifications shown in Annexure IV below:-

S.No.	Item Name	Technical Specification	Qty.
1.	Vacuum Manifold with flask (6 Places each)	Annexure IV	2

2. The tenders should be submitted in two sealed covers:-

(a) The first sealed cover superscribed as 'Technical Bid' should contain the following items:-

(i) The Performa at Annexure -II, duly filled in, along with relevant documents/information.

(ii) Acceptance of terms and conditions as at Annexure - I


(b) The second sealed envelope superscribed 'Financial Bid' as at Annexure-III should contain only rates for supply of Vacuum Manifold with flask (6 Places each) .

(c) Both the sealed covers should be placed in the main sealed envelope super scribed 'Tender for supply of Vacuum Manifold with flask (6 Places each). Tender should be addressed to Director Regional Drugs Testing Laboratory, Sector 39-C, Chandigarh and should reach this office by 3.30. P.M. on 07.09.2018.

3. Quotation received after stipulated time & date will be rejected forthwith.

4. The competent authority in the RDTL reserves the right to amend any of the term and conditions in the tender document or reject any or all application (offers) without giving any notice or assigning any reason thereof. The decision of competent authority in this regard will be final and binding.

5. All the tenderers are requested to read and understand the terms and conditions of the contract as detailed in the Annexure - I before sending their quotations, as no change or violation of the aforesaid terms and conditions are permissible once the quotation is accepted by RDTL.


Dr. R.A.Singh
Director

Terms and Conditions

1. Necessary technical literature and catalogue should be forwarded along with Tender document.
2. Rates quoted for the supply of Vacuum Manifold with flask (6 Places each) should be inclusive of supply/freight/taxes etc. and no extra payment will be made thereafter. Further, the number of the system may increase or decrease as per requirement of the competent authority.
3. Only manufacturing firms/authorized distributors are eligible to participate. While Submitting the Tenders, the tendering firm shall have to furnish the proof of PAN No., Service Tax No., Sales Tax No., etc. as per Annexure-II
4. The tender received after due date and time or incomplete in any respect shall be summarily rejected.
5. No Advance payment would be made. Payment will be made only after complete supply and satisfactory working of the system.
6. **The shortlist firm will have to provide the warranty for at least 2 years.**
7. Over-writing on the quotations must be avoided. Otherwise the quotations are likely to be rejected. The RDTL reserves the right to accept or reject any tender, in whole or in part thereof without assigning/ specifying any reason (s) thereof and the decision of the competent authority in the RDTL shall be final and binding on the contracting firm.
10. Any dispute arising in the matter shall be resolved through an arbitrator to be nominated by the competent authority in RDTL.
11. The contract shall be subject to Indian Laws and jurisdiction of the court located in Chandigarh.
12. The system quoted should be latest model and spares/ services should be available for next 10 years.
13. Instrument offered should be new and latest model and refurbished or old equipment will no be accepted.

Annexure-II

PROFORMA (Technical bid)

S.No.	Particulars	To be filled by the Tenderer
1.	Name of the agency	
2.	Whether brief profile of the agency is enclosed	
3.	Detailed Office Address of the agency with office telephone No. e-mail address, Fax No., Mobile No. and name/designation of the contract person.	
4.	Whether registered with all concerned Govt. authorities (Register of companies, Commissioner Employees etc.)(Copies of all certificates of Registration should be attached)	
5.	PAN Number (copies to be enclosed)	
6.	TIN Number Registration No. (Copy to be enclosed)	
7.	Service Tax Registration No. (copy to be enclosed)	
10.	Whether copies of income Tax return for last 2 years attached	
11.	Whether a copy of terms and conditions- Annexure-I duly signed as token of acceptance of the' same is attached	
12.	Whether the firm is blacklisted by any Government Department or any criminal case is registered against the firm or its owner/partners anywhere in India (if no, an undertaking to this effect is attached in this regard.)	

(Signature of the authorized person)

Date:
Place:

Name:
Designation:
Seal:

Annexure-III

Financial Bid for Supply of Vacuum Manifold with flask (6 Places each) in RDTL, Sector 39-C, Chandigarh.

With reference to RDTL letter No.....dated.....Regarding tender for supply of Vacuum Manifold with flask(6 places each), I/we quote the rate, inclusive of supply/installation/freight etc. As under:-

S.No	Item Name/Qty	Specifications	Cost of Apparatus
1.	Vacuum Manifold with flask (6 Places each)		

IMPORTANT:

In case of discrepancy between the unit price and total price, the unit price shall prevail for computation.

Date:

Place:

(Signature of the authorized person)

Name:

Designation:

Seal:

Specification of Manifold with flask:

Number of Branches	:	6
Material	:	Stainless steel 316
Valve Type	:	2 or 3 way
Control	:	Individual control of vacuum with suitable valve.
Pressure	:	Vacuum only.
Sterilization	:	Autoclavable by moist heat sterilization.
Fitting	:	Proper leak proof fittings with below mentioned flasks
Outlet Connection	:	stainless steel hose connector and pipe plug, O.D. ¼ inch (6.4 mm).

Flask with clamps and support screen:

Material (Funnel, base and lid):	:	Stainless steel 316.
Capacity of flask	:	500 ml
Dia. of filters to be used	:	47 mm
Pressure	:	Vaccum
Sterilization	:	Autoclavable by moist heat sterilization.
Locking system	:	Turn lock adaptor for proper sealed locking.
Warranty	:	Minimum 2 years from the date of supply.

The certificates of Stainless steel 316 material have to be provided