

भारतसरकार
क्षेत्रीय औषध परीक्षणप्रयोगशाला

केन्द्रीय औषध मानकनियन्त्रण संगठन
(स्वास्थ्य सेवाओंका महानिदेशालय)
स्वास्थ्य एवंपरिवारकल्याणमंत्रालय
सैक्टर-39 सी, चन्डीगढ़-160036
फोननं: 0172-2688239, फैक्सनं: 0172-2636316



NABL

ACCREDITATED LABORATORY

GOVERNMENT OF INDIA
REGIONAL DRUGS TESTING LABORATORY

CENTRAL DRUGS STANDARD CONTROL ORGANISATION
(Directorate General of Health services)

Ministry of Health & Family Welfare

SECTOR-39C, CHANDIGARH – 160036

Tel. No: -0172-2688239, Fax- No: 0172-2636316

E-mail id: rdtlchd@cdsco.nic.in

Tender Notice

Subject: Tender Rate Enquiries for Supply of RP HPLC Column with Compatible Guard Column/Cartridge with Holder at RDTL, Chandigarh

Sealed tenders are invited from interested manufacturers, dealers, distributors, or retailers for the supply of RP HPLC Column along with compatible guard column/cartridge and holder (where required) at RDTL, Chandigarh, as specified in **Annexure-A**.

1. Tenders should be submitted in a sealed envelope containing;
 - (i) The Proforma at **Annexure-II** and **Annexure-III**, duly filled along with relevant documents/information.
 - (ii) Acceptance of terms and conditions as per **Annexure-I**.
2. The sealed envelope should be superscribed with “Tender Rate Enquiry for Supply of RP HPLC Column with Compatible Guard Column/Cartridge with Holder (wherever required).” Tenders should be addressed to the Director I/c, Regional Drugs Testing Laboratory, Sector 39-C, Chandigarh, and must reach this office within 21 days from the date of publishing of this tender, i.e., by 3:00 PM on 01.08.2024.
3. The quotations received will be opened on the same day, i.e., on 01.08.2024 at 3:30 PM in the presence of interested bidders or their authorized representatives.
4. Tenders received after the stipulated date and time will be rejected forthwith.
5. The competent authority at RDTL reserves the right to amend any of the terms and conditions in the tender document, reject any or all applications (offers), or cancel the tender without notice or reason. The decision of the competent authority in this regard will be final and binding.

Director I/c

Annexure-I**Terms and Conditions**

1. **Rates and Payment:** The rates quoted for the RP HPLC Column with compatible guard column/cartridge with holder (wherever required) should include supply, freight, taxes, etc. No extra or advance payment will be made for the supply of the required items. Payment will be made only after the receipt of materials in good condition and satisfactory performance test. Additionally, the quantity of items may increase or decrease as per the department's requirements.
2. **Eligible Brands and Documentation:** Only the brands Supelco, Merck, Agilent, Hichrom, Waters, Diacel, SGE, Hamilton, and Thermo Fisher Scientific are eligible to participate in the tender. Any other brand will not be considered. When submitting tenders, the tendering firm must furnish proof of PAN No., GSTIN, and Authorization of the Manufacturer in the case of dealers, as per Annexure-II.
3. **Technical Specifications:** The part number of the quoted material must be clearly mentioned to verify the online technical specifications on the original manufacturer's website. Additionally, the online link or web address must be provided. The components must strictly meet the technical specifications as outlined in Annexure-A. Deviations from these specifications will lead to rejection of the tender.
4. **Quality Assurance:** The bidder must provide a commitment to ensure that all materials supplied are of high quality and comply with industry standards and specifications mentioned in Annexure-A.
5. **Compatibility Requirement:** The quoted columns must be compatible with the Thermo Fisher Ultimate 3000 HPLC without any extra adapters.
6. **Submission Requirements:** Tenders received after the stipulated time or incomplete in any respect will be summarily rejected.
7. **Delivery:** The supply of items must be made within 30 days from the date of issue of the purchase order; otherwise, the purchase order will be canceled.
8. **Manufacturing Date and Traceability:** The date of manufacturing of the RP HPLC Column with compatible guard column/cartridge with holder (wherever required) should not be more than one year old. All products should be traceable online on the manufacturer's website using the serial number of the supplied material; otherwise, the supplied items will be rejected.
9. **Quotation Requirements:** Overwriting on the quotations must be avoided; otherwise, the quotations are likely to be rejected. The RDTL reserves the right to accept or reject any tender, in whole or in part, without assigning any reason, and the decision of the competent authority in the RDTL shall be final and binding on the bidder.
10. **Dispute Resolution:** Any dispute arising in the matter shall be resolved through arbitration nominated by the competent authority in RDTL.
11. **Jurisdiction:** The contract shall be subject to Indian laws and the jurisdiction of the courts located in Chandigarh.
12. **Understanding of Terms and Conditions:** All bidders are requested to thoroughly read and understand the terms and conditions of the contract as detailed in Annexure-I before submitting their quotations. No changes or violations of these terms and conditions are permissible once the quotation is accepted by RDTL, Chandigarh.
13. **GeM Unique ID:** GeM availability report and past transaction summary unique ID: GEM/GARPTS/04062024/TA8TR9KUPNHF.

Annexure-II**PROFORMA (Tender bid)**

S.No.	Particulars	To be filled by the Tenderer
1.	Name of the Agency/Firm	
2.	Whether brief profile of the agency is enclosed	
3.	Detailed Office Address of the agency with office telephone No. e-mail address, Fax No., Mobile No. and name/designation of the contract person.	
4.	PAN Number (copies to be enclosed)	
5.	TIN Number Registration No. (Copy to be enclosed)	
6.	Service Tax Registration No. (copy to be enclosed)	
7.	Whether a copy of terms and conditions- Annexure-I duly signed as token of acceptance of the' same is attached	
8.	Whether the firm is blacklisted by any Government Department or any criminal case is registered against the firm or its owner/partners anywhere in India (if no, an undertaking to this effect is attached in this regard.)	

(Signature of the authorized person)

Date:

Place:

Name:

Designation:

Seal:

Annexure-III

FINANCIAL BID FOR SUPPLY OF RP HPLC COLUMN WITH COMPATIBLE GUARD COLUMN/CARTRIDGE WITH HOLDER (WHEREVER IS REQUIRED) IN RDTL Sector 39-C, Chandigarh.

With reference to RDTL letter No.....dated.....Regarding tender for supply of RP HPLC COLUMN WITH COMPATIBLE GUARD COLUMN/CARTRIDGE WITH HOLDER (WHEREVER IS REQUIRED), I/we quote the rate, inclusive of supply/ freight etc. As under:-

Sr.No	Item Name/Qty.	Quantity	Cost of item
	As per Annexure- A attached		
	Taxes, if any		
	Grand Total		

IMPORTANT:

In case of discrepancy between the unit price and total price, the unit price shall prevail for computation.

Date:

Place:

(Signature of the authorized person)

Name:

Designation:

Seal:

ANNEXURE-A

Sl.No.	USP Type	Technical Specification	Quantity
1. C-18 BDS (USP Type L1)			
1. (a)	C-18 BDS (USP Type L1)	Column Format – Analytical Column Length (mm) – 250 ID (mm) – 4.6 Particle Size- 5 µm pH – 2 to 9 or better Pore Size (Å) – 100 or less Surface Area (m ² /g) – 320 or higher Carbon load – 16 to 18% Packing Material – Spherical, Fully Porous, Base- Deactivated Silica/ Ultrapure Silica, BEH C18 Endcapped- Yes Column Type- Reversed Phase Pressure- 400 bar (5800 psi) or higher	06 Nos.
1. (b)	C-18 BDS (USP Type L1)	Compatible guard column/cartridge with holder(whenever is required) for the HPLC Columns, mentioned in 1.(a)	20 Nos.