

भारतसरकार  
क्षेत्रीय औषध परीक्षणप्रयोगशाला

केन्द्रीय औषध मानकनियन्त्रण संगठन  
(स्वास्थ्य सेवाओंका महानिदेशालय)  
स्वास्थ्य एवंपरिवारकल्याणमंत्रालय  
सैक्टर-39 सी, चन्डीगढ़-160036  
फोननं: 0172-2688239, फैक्सनं: 0172-2636316



**NABL**

ACCREDITATED LABORATORY

GOVERNMENT OF INDIA  
REGIONAL DRUGS TESTING LABORATORY

CENTRAL DRUGS STANDARD CONTROL ORGANISATION  
(Directorate General of Health services)

Ministry of Health & Family Welfare

SECTOR-39C, CHANDIGARH – 160036

Tel. No: -0172-2688239, Fax- No: 0172-2636316

E-mail id: [rdtlchd@cdsco.nic.in](mailto:rdtlchd@cdsco.nic.in)

**Tender Notice**

**Subject: Tender Rate Enquiries for Supply of 02 nos. ATC Probes and 01 nos. sealed, double-junction, open-pore pH electrode with BNC Compatible with pH meters, Make: Mettler Toledo, Model: S210 at RDTL, Chandigarh.**

Sealed tenders are invited from interested manufacturers, dealers, distributors, or retailers for the supply of Supply of 02 nos. ATC Probes and 01 nos. sealed, double-junction, open-pore pH electrode with BNC Compatible with pH meters, Make: Mettler Toledo, Model: S210 as specified in Annexure-A.

1. Tenders should be submitted in a sealed envelope containing;
  - (i) The Proforma at **Annexure-II** and **Annexure-III**, duly filled along with relevant documents/information.
  - (ii) Acceptance of terms and conditions as per **Annexure-I**.
2. The sealed envelope should be superscribed with “Tender Rate Enquiry for 02 nos. ATC Probes and 01 nos. sealed, double-junction, open-pore pH electrode with BNC Compatible with pH meters, Make: Mettler Toledo, Model: S210 at RDTL, Chandigarh.” Tenders should be addressed to the Director I/c, Regional Drugs Testing Laboratory, Sector 39-C, Chandigarh, and must reach this office within 21 days from the date of publishing of this tender, i.e., by 3:00 PM on 20.09.2024.
3. The quotations received will be opened on the same day, i.e., on 20.09.2024 at 3:30 PM in the presence of interested bidders or their authorized representatives.
4. Tenders received after the stipulated date and time will be rejected forthwith.
5. The competent authority at RDTL reserves the right to amend any of the terms and conditions in the tender document, reject any or all applications (offers), or cancel the tender without notice or reason. The decision of the competent authority in this regard will be final and binding.

**Director I/c**

**Annexure-I****Terms and Conditions**

1. **Rates and Payment:** The rates quoted for 02 nos. ATC Probes and 01 nos. sealed, double-junction, open-pore pH electrode with BNC Compatible with pH meters, make: Mettler Toledo, Model: S210 should include supply, freight, taxes, etc. No extra or advance payment will be made for the supply of the required items. Payment will be made only after the receipt of materials in good condition and satisfactory performance test. Additionally, the quantity of items may increase or decrease as per the department's requirements.
2. **Documentation:** The tendering firm must furnish proof of PAN No., and GSTIN as per Annexure-II.
3. **Technical Specifications:** The part number of the quoted material must be clearly mentioned to verify the online technical specifications on the original manufacturer's website. Additionally, the online link or web address must be provided. The components must strictly meet the technical specifications as outlined in Annexure-A. Deviations from these specifications will lead to rejection of the tender.
4. **Quality Assurance:** The bidder must provide a commitment to ensure that all materials supplied are of high quality and comply with industry standards and specifications mentioned in Annexure-A.
5. **Compatibility Requirement:** The quoted items must be compatible with the pH Meter of Mettler Toledo, Model S210.
6. **Submission Requirements:** Tenders received after the stipulated time or incomplete in any respect will be summarily rejected.
7. **Delivery:** The supply of items must be made within 30 days from the date of issue of the purchase order; otherwise, the purchase order will be canceled.
8. **Quotation Requirements:** Overwriting on the quotations must be avoided; otherwise, the quotations are likely to be rejected. The RDTL reserves the right to accept or reject any tender, in whole or in part, without assigning any reason, and the decision of the competent authority in the RDTL shall be final and binding on the bidder.
9. **Dispute Resolution:** Any dispute arising in the matter shall be resolved through arbitration nominated by the competent authority in RDTL.
10. **Jurisdiction:** The contract shall be subject to Indian laws and the jurisdiction of the courts located in Chandigarh.
11. **Understanding of Terms and Conditions:** All bidders are requested to thoroughly read and understand the terms and conditions of the contract as detailed in Annexure-I before submitting their quotations. No changes or violations of these terms and conditions are permissible once the quotation is accepted by RDTL, Chandigarh.
12. **GeM Unique ID:** GeM availability report and past transaction summary unique ID: GEM/GARPTS/23082024/5Y7ZBOC9JANZ and : GEM/GARPTS/23082024/WJX6VSBQ4QU5.

**Annexure-II****PROFORMA (Tender bid)**

S.No.	Particulars	To be filled by the Tenderer
1.	Name of the Agency/Firm	
2.	Whether brief profile of the agency is enclosed	
3.	Detailed Office Address of the agency with office telephone No. e-mail address, Fax No., Mobile No. and name/designation of the contract person.	
4.	PAN Number (copies to be enclosed)	
5.	TIN Number Registration No. (Copy to be enclosed)	
6.	Service Tax Registration No. (copy to be enclosed)	
7.	Whether a copy of terms and conditions- Annexure-I duly signed as token of acceptance of the' same is attached	

(Signature of the authorized person)

Date:  
Place:

Name:  
Designation:  
Seal:

**Annexure-III****FINANCIAL BID FOR SUPPLY OF 02 NOS. ATC PROBES AND 01 NOS. SEALED, DOUBLE-JUNCTION, OPEN-PORE PH ELECTRODE WITH BNC COMPATIBLE WITH PH METERS, MAKE: METTLER TOLEDO, MODEL: S210 IN RDTL SECTOR 39-C, CHANDIGARH.**

With reference to RDTL letter No.....dated.....Regarding tender for supply of 02 NOS. ATC PROBES AND 01 NOS. SEALED, DOUBLE-JUNCTION, OPEN-PORE PH ELECTRODE WITH BNC COMPATIBLE WITH PH METERS, MAKE: METTLER TOLEDO, MODEL: S210, I/we quote the rate, inclusive of supply/ freight etc. As under:-

Sr.No	Item Name/Qty.	Quantity	Cost of item
	As per Annexure- A attached		
	Taxes, if any		
	Grand Total		

**IMPORTANT:**

In case of discrepancy between the unit price and total price, the unit price shall prevail for computation.

Date:

Place:

(Signature of the authorized person)

Name:

Designation:

Seal:

**ANNEXURE-A**

<b>Sr.No.</b>	<b>Technical Specifications of ATC Probe</b> <b>(Compatible with S210 pH METERS, MAKE: METTLER TOLEDO, MODEL: S210)</b>	
1.	Measuring range	30 °C – 100 °C
2.	Accessory Type	Temperature probes for pH meter
3.	Cable and connectors	1.0 m, RCA (Cinch)
4.	Intelligent Sensor Management (ISM)	No
5.	Temperature Probe	NTC 30 kΩ
<b>Sr.no.</b>	<b>Technical Specification of pH electrode</b> <b>(Compatible with S210 pH METERS, MAKE: METTLER TOLEDO, MODEL: S210)</b>	
1.	SEALED, DOUBLE-JUNCTION, OPEN-PORE PH ELECTRODE WITH BNC connector	