

No. 6-4/2018-S/ 47
OFFICE OF THE DIRECTOR
CENTRAL DRUGS LABORATORY,
GOVERNMENT OF INDIA
MIN. OF HEALTH & FAMILY WELFARE
3, KYD STREET, KOLKATA-16

Kolkata Dated the 07/05/2018

NOTICE INVITING QUOTATION

The Director-in Charge, Central Drugs Laboratory, Kolkata invites quotation from the interested Manufacturer, Distributer or Dealer for providing and supplying of **Bacteriological Incubator** at Central Drugs Laboratory, Kolkata.

For detailed Notice. please visit <http://www.cdscn.co.in>

The detailed terms & condition of quotation for **Bacteriological Incubator** is available on the above website. Filled in quotation, accepting the stipulated terms & condition may be dropped in tender box available at the wing of Jr. Administrative Officer, Central Drugs Laboratory, Kolkata. Tender / Quotation on any working day between 9:30 am. to 6:00 pm. and upto 3:00 pm **on 14.06.2018**. The quotation will be opened on the same day **at 3:30 pm.** in presence of interested Quotationers.

Yours faithfully


7/5/18

Dr. Anjan Pal,
SSO (Microbiology)
for Director

Central Drugs Laboratory, Kolkata

**No. 6-4/2018-S/
OFFICE OF THE DIRECTOR
CENTRAL DRUGS LABORATORY,
GOVERNMENT OF INDIA
MIN. OF HEALTH & FAMILY WELFARE
3, KYD STREET, KOLKATA-16**

Kolkata Dated the **07/05/2018**

NOTICE INVITING QUOTATION

Quotation is hereby invited from the interested and bona fide Dealer, Distributers, Retailers for Providing and supplying of **Bacteriological Incubator** under the following specification, terms and conditions:


Name of the Item	Quantity
<u>Bacteriological Incubator</u> -	2 Nos.
<u>Specification :</u>	
1. Temperature Range : Ambient +5°C to 60°C	
2. Temperature Control Accuracy: ±0.2°C of set point or better.	
3. Temperature Uniformity : ±0.5°C to ±0.7°C	
4. Control Type : Microprocessor	
5. Temperature Display : LED	
6. Temperature Sensor : PT 100	
7. Capacity 40 to 60 liter, shelves-2	
8. With motorized fan blower for air circulation	
9. Heating U-shaped Nichrome wire in SS sheathing	
10. Double walled with insulation provided with outer stainless steel metal door and inner glass viewing door.	
11. Input Voltage : 230 volts AC, 50Hz.	
12. Outer cabinet SS 304 grade dull matt finish	
13. Trays SS wire mesh, heavy duty	
14. Inner chamber stainless steel mirror finish	
15. With PQ, IQ, OQ documentation	
16. Digital timer of 999 minutes with auto heater cut off for timed cycle.	
17. High temperature safety cut off with audio / visual alarm.	
18. Performance qualification with temperature mapping on site.	
19. Optional Accessories : Printer interface facility to record data of 1 or 3 chambers.	
20. Optional :Magnetic door lock feature	
21. Onsite validation for loaded condition and unloaded condition to be done by the supplier on installation and there after every six months during warranty period.	

N.B: Applicable Taxes and duties if any, which may be levied by the Government, should be shown separately.

Terms & Conditions. :

1. The quotation/tender should be submitted in sealed covers. Open quotation/tender will be summarily rejected.
2. The quotation should be valid for a period of 90 (ninety) days from the date of opening of the tender / quotation.
3. The Director-in-Charge, CDL, Kolkata reserves the right to reject the quotation and/or supply order at any stage without assigning any reasons thereof.
4. The quoted rate should be all inclusive and nothing will be paid extra for transportation, loading, unloading at Central Drugs Laboratory, Kolkata.
5. The Quotationer/supplier should supply of **Bacteriological Incubator** in good condition & strictly as per specification within thirty (30) days from the date of issue of supply order.
6. The person duly authorized by the quotationer will be allowed to deliver **Bacteriological Incubator** at Central Drugs Laboratory, Kolkata on any working day within the time schedule.
7. No "Way Bill" in connection with supply of **Bacteriological Incubator** will be issued from Central Drugs Laboratory, Kolkata.
8. **The quotationer should clearly mention the name of manufacturer and Product No.**
9. That on successful completion of supply of **Bacteriological Incubator** as per specification, make and Catalogue Number to Store Section, CDL, Kolkata, the quotationer/supplier should submit bill in triplicate along with mandate form to release e-payment through Pay & Accounts Office, Min. of Health & F.W. Kolkata.
10. **Only manufacturing firms / authorized distributors are eligible to participate.**
11. The Quotationer should clearly quote their rates for both in figures and in words **in net price (including any other taxes/discounts etc.) failing which your quotation will not be entertained.** Any overwriting in rates or any change and/or alteration without attestation is liable to culminate in the outright rejection of the quotation. In case there is any error between the amount specified in figures and words, the lower of the two will be taken for consideration.
12. The Quotation duly completed as per the instruction given herein and accepting the same may be dropped in the "Tender Box" on any working days between **11.00 AM to 3.00 P.M. up to 14.06.2018** at the wing of Jr. Administrative Officer, Central Drugs Laboratory, Government of India, Ministry of Health & Family Welfare, 3, Kyd Street, Kolkata-700016.
13. Late quotation i.e. quotation received after the specified date and time of receipt will not be considered in any circumstances.
14. The Quotation so received will be opened on the same day i.e. on **14.06.2018 at 3.30 P.M.** in presence of interested Quotationer(s) or their authorized representative.

Yours faithfully



Dr. Anjan Pal,
SSO (Microbiology)
for Director

Central Drugs Laboratory, Kolkata.