

No. 6-4(Store)/2017-18/ 2369  
OFFICE OF THE DIRECTOR  
CENTRAL DRUGS LABORATORY,  
GOVERNMENT OF INDIA  
MIN. OF HEALTH & FAMILY WELFARE  
3, KYD STREET, KOLKATA-16

Kolkata Dated the 21-09-17


**NOTICE INVITING QUOTATION**

The Director-in Charge, Central Drugs Laboratory, Kolkata invites quotation from the interested and bonafide Manufacturer, Dealer, Distributer or retailer providing and supplying of **Table Top Horizontal Airflow Workstation** at Central Drugs Laboratory, Kolkata.

**For detailed Notice. please visit <http://www.cdscsco.nic.in>**

The detailed terms & condition of quotation for providing and supplying of **Table Top Horizontal Airflow Workstation** is available on the above website. Filled in quotation, accepting the stipulated terms & condition may be dropped in tender box available at the wing of Jr. Administrative Officer, Central Drugs Laboratory, Kolkata on any working day upto **3:00pm on 26.10.2017** and Tender / Quotation will be opened on the same day at **3:30pm on 26.10.2017** in presence of interested Quotationers.

Yours faithfully

  
(Dr. Anjan Pal) 21/9/17  
SSO (Microbiology)  
for Director  
Central Drugs Laboratory, Kolkata.

Kolkata Dated the


**NOTICE INVITING QUOTATION**

Quotation is hereby invited from the interested and bona fide Dealer, Distributors, Retailers for Providing and supplying of **Table Top Horizontal Airflow Workstation** under the following specification, terms and conditions:

Specification	Description
Work Area (WxDxH)	Inner Dimension: 762mm x 610mm x 762mm (approx). Overall Height : ≤5 feet without stand.
Air Filter Type	Two stage Filtration : <b>Pre-filter</b> : Should be made of anti-microbial <b>EU6/F6</b> grade <b>mini pleat</b> or equivalent type water repellent filter media with an efficiency of 60-80% (approximately 95% down to <b>3 micron</b> ) or better. <b>HEPA Filter</b> : Should be made of Anti-microbial <b>EU13/H14</b> grade <b>mini pleat</b> or equivalent type water repellent HEPA Filter with an efficiency of 99.97% (approximately 99.99% down to <b>3 micron</b> ) or better.
Air Flow	0.45m±0.05m/sec or higher with ON / OFF switch
Air Cleanliness	ISO Class 5 or Class <b>M3.5</b> or <b>Class 100</b> or equivalent
Air Flow Direction	<b>Horizontal</b>
Pressure Monitor	Should maintain 0-25mm differential pressure across HEPA Filter with respect to ambient
Work Area	Should be made of <b>SS304</b>
Lamp	a) Fluorescent lamp (in excess of <b>200 lux</b> ) with ON / OFF Switch b) 1x8W (approx) with minimum vibration
Power Supply	230 VAC and 2x5/15 Amp electrical socket with ON / OFF switch for external equipment with CE compliance
Noise Level	≤70dB (approx) with minimum vibration
Door and Side Panels	Made of transparent, UV protecting and chemical resistant materials. Fully openable front door with gas filled hinges
Motor Blower Position	<b>Top mounted</b>

**N.B:** Applicable Taxes and duties if any, which may be levied by the Government, should be shown separately.

Yours faithfully

  
Dr. Anjan Pal,  
SSO (Microbiology)  
for Director

Central Drugs Laboratory, Kolkata.

**Terms & Conditions. :**

1. The quotation/tender should be submitted in sealed covers. Open quotation/tender will be summarily rejected.
2. The quotation should be valid for a period of 90 (ninety) days from the date of opening of the tender / quotation.
3. The Director-in-Charge, CDL, Kolkata reserves the right to reject the quotation and/or supply order at any stage without assigning any reasons thereof.
4. The quoted rate should be all inclusive and nothing will be paid extra for transportation, loading, unloading at Central Drugs Laboratory, Kolkata.
5. The Quotationer/supplier should supply of **Table Top Horizontal Airflow Workstation** in good condition & strictly as per specification within thirty (30) days from the date of issue of supply order.
6. The person duly authorized by the quotationer will be allowed to deliver **Table Top Horizontal Airflow Workstation** at Central Drugs Laboratory, Kolkata on any working day within the time schedule.
7. No "Way Bill" in connection with supply of **Table Top Horizontal Airflow Workstation** will be issued from Central Drugs Laboratory, Kolkata.
8. **The quotationer should clearly mention the name of manufacturer and Product No.**
9. That on successful completion of supply of **Table Top Horizontal Airflow Workstation** as per specification, make and Catalogue Number to Store Section, CDL, Kolkata, the quotationer/supplier should submit bill in triplicate along with mandate form to release e-payment through Pay & Accounts Office, Min. of Health & F.W. Kolkata.
10. **Only manufacturing firms / authorized distributors are eligible to participate.**
11. The Quotationer should clearly quote their rates for both in figures and in words **in net price (including any other taxes/discounts etc.) failing which your quotation will not be entertained.** Any overwriting in rates or any change and/or alteration without attestation is liable to culminate in the outright rejection of the quotation. In case there is any error between the amount specified in figures and words, the lower of the two will be taken for consideration.
12. The Quotation duly completed as per the instruction given herein and accepting the same may be dropped in the "Tender Box" on any working days between **11.00 AM to 3.00 P.M. up to 26.10.2017** at the wing of Jr. Administrative Officer, Central Drugs Laboratory, Government of India, Ministry of Health & Family Welfare, 3, Kyd Street, Kolkata-700016.
13. Late quotation i.e. quotation received after the specified date and time of receipt will not be considered in any circumstances.
14. The Quotation so received will be opened on the same day i.e. on **26.10.2017 at 3.30 P.M.** in presence of interested Quotationer(s) or their authorized representative.

Yours faithfully



Dr. Anjan Pal, 21/9/17  
SSO (Microbiology)  
for Director

Central Drugs Laboratory, Kolkata.