

भारत सरकार
क्षेत्रीय औषध परीक्षण प्रयोगशाला
केन्द्रीय औषध मानक नियन्त्रण संगठन
(स्वास्थ्य सेवाओं का महानिदेशालय)
स्वास्थ्य एवं परिवार कल्याण मंत्रालय
सेक्टर-39 सी, चण्डीगढ़-160036
फोन नं: 0172-2688239, फैक्स नं: 0172-2636316



NABL

ACCREDITED LABORATORY

GOVERNMENT OF INDIA
REGIONAL DRUGS TESTING LABORATORY

CENTRAL DRUGS STANDARD CONTROL ORGANISATION
(Directorate General of Health services)
Ministry of Health & Family Welfare
SECTOR-39 C, CHANDIGARH – 160036
Tel. No: -0172-2688239, Fax- No: 0172-2636316
E-mail id: rdtlchd@cdsco.nic.in

No. 3-18/2021-22/RDTL/St./Part-I/2689 /21

Dated: 29.12.2021

Tender notice

Subject:-Tender Rate Enquiry for supply of PRTs (Platinum resistant thermometers) in RDTL, Chandigarh.

Sealed Tender Rate Enquiries are invited from reputed interested and manufacturing firm(s)/distributors of PRTs (Platinum resistant thermometers) in RDTL, who have been rendering/undertaking the specified work in the field of Scientific test equipment as per the specifications shown in Annexure IV below:-

S.No.	Item Name	Technical Specification	Qty.
1.	Platinum resistant thermometers	Annexure IV	3

- The tenders should be submitted in sealed envelope containing;
 - The Performa at **Annexure-II** and **Annexure-III**, duly filled in along with relevant documents/information.
 - Acceptance of terms and conditions as per **Annexure – I**.
- The sealed envelope should be super scribed with "Tender Rate Enquiry for supply of Platinum resistant thermometers". Tender should be addressed to the Director, Regional Drugs Testing Laboratory, Sector 39-C, Chandigarh and be reached in this office within 21 days from the date of publishing of this tender, i.e. 19.01.2022 and at 3:00 PM.
- The quotations, so received will be opened on same day i.e. on 19.01.2022 at 3:30 PM in presence of interested quotationer(s) or their authorized representatives.
- Tender received after stipulated date & time will be rejected forthwith.
- The competent authority in the RDTL reserves the right to amend any of the terms and conditions in the tender document or reject any or all applications (offers) or cancel the tender without giving any notice or assigning any reason thereof. The decision of competent authority in this regard will be final and binding.

Dr. R.A.Singh
Director

Terms and Conditions

1. Necessary technical literature and catalogue should be forwarded along with Tender document.
2. Rates quoted for Platinum resistant thermometers should be inclusive of supply/freight etc. and no extra payment will be made for supply of above items. Further, the number of the item may be changed according to the requirement of the Department.
3. Only manufacturing firms/authorized distributors of Platinum resistant thermometers are eligible to participate. While Submitting the Tenders, the tendering firm shall have to furnish the proof of PAN No., GST, etc. as per Annexure-II
4. The quotation duly filled in as per the instruction giving herein may be submitted directly or by post to this office up to 3:00 PM on dated: 19.01.2022.
5. The tender received after due date and time or incomplete in any respect shall be summarily rejected.
6. No Advance payment would be made. Payment will be made only after complete supply and satisfactory working of the equipment.
7. **The shortlist firm will have to provide the warranty for at least 2 years.**
8. Over-writing on the quotations must be avoided; Otherwise, the quotations are likely to be rejected. The competent authority of this laboratory reserves the right to accept or reject any tender, in whole or in part without assigning/ specifying any reason (s) thereof and the decision of the competent authority in the RDTL shall be final and binding on the contracting firm.
9. All the participating firms should ensure that the equipments are of only Genuine, good quality and in good condition, if at any stage it is found that the item supplied is of sub-standard quality/inferior quality will be rejected forth with and the firm will be blacklisted.
10. If in any circumstances, the first lowest successful bidder refuses or being unable to fulfill his obligations within the stipulated time frame, then the second successful lowest bidder will be offered and so on up to the lowest bidder where the price can be compared with at least 3 other bidder's prices.
11. Any dispute arising in the matter shall be resolved through an arbitrator to be nominated by the competent authority in RDTL.
12. The contract shall be subject to Indian Laws and jurisdiction of the court located in Chandigarh.
13. The system quoted should be latest model and spares/ services should be available for next 10 years.
14. Instrument offered should be new and latest model and refurbished or old equipment will not be accepted.
15. The calibration certificates to be provided with the PRT at the time of delivery.

Annexure-II

PROFORMA (Technical bid)

S.No.	Particulars	To be filled by the Tenderer
1.	Name of the agency	
2.	Whether brief profile of the agency is enclosed	
3.	Detailed Office Address of the agency with office telephone No. e-mail address, Fax No., Mobile No. and name/designation of the contract person.	
4.	Whether registered with all concerned Govt. authorities (Register of companies, Commissioner Employees etc.)(Copies of all certificates of Registration should be attached)	
5.	PAN Number (copies to be enclosed)	
6.	TIN Number Registration No. (Copy to be enclosed)	
7.	Service Tax Registration No. (copy to be enclosed)	
10.	Whether copies of income Tax return for last 2 years attached	
11.	Whether a copy of terms and conditions- Annexure-I duly signed as token of acceptance of the' same is attached	
12.	Whether the firm is blacklisted by any Government Department or any criminal case is registered against the firm or its owner/partners anywhere in India (if no, an undertaking to this effect is attached in this regard.)	

(Signature of the authorized person)

Date:

Place:

Name:

Designation:

Seal:

Annexure-III

Financial Bid for Supply of Platinum resistant thermometers in RDTL, Sector 39-C, Chandigarh.

With reference to RDTL letter No.....dated.....Regarding tender for supply of Platinum resistant thermometers, I/we quote the rate, inclusive of supply/installation/freight etc. As under:-

S.No	Item Name/Qty	Specifications	Cost of Platinum resistant thermometers
1.	Platinum resistant thermometers		

IMPORTANT:

In case of discrepancy between the unit price and total price, the unit price shall prevail for computation.

Date:

Place:

(Signature of the authorized person)

Name:

Designation:

Seal:

Technical Specifications of Platinum Resistant Thermometer

S.No	Specification	Platinum Resistant Thermometer	Platinum Resistant Thermometer
01	Temperature Range	-80 °C to 400 °C or Better	0 °C to 1000 °C or Better
02	Temperature accuracy	±1 °C or better (up to 400 °C)	±5 °C or better (at 1000 °C)
03	Sheath material	SS-316	SS-316
04	Dimensions	Standard Dimensions with long sensor and 2 meter long well insulated cable	Standard Dimensions with long sensor and 2 meter long well insulated cable
05	Handle dimension	Standard dimensions for suitable handling	Standard dimensions for suitable handling
06	Display	Digital	Digital
07	Calibration	As per NABL standards	As per NABL standards
	Qty.	01	02